

CERTIFICATE OF TESTING No: 6535/09

Issue 1

Customer: Bioquell UK Ltd
53 Royce Close
Portway Ind Estate
Andover
Hants, SP10 3TS

Customer Reference No: STD224259 TRaC ID No: 25-0019-5422

Equipment Tested

EUT: Clarus L2 Model No: TD059-0000-230V Serial No: 2009030306

Test Specifications

EN 61326:2006

Date of Test: 9th-11th March 2009 Test Location: TRaC Ringwood

Comments: See summary on page 2

Test Results: COMPLIANT

This is to certify that the above referenced equipment has been tested in accordance with the requirements of the contract/order/specification as detailed above using test equipment calibrated to traceable National Standards. This Certificate applies only to the product(s) detailed and is not indicative of the qualities of identical or similar products.

Approved By:



S Youngman (General Manager)

Date of Issue: 01/04/2009

Page 1 of 2

The results herein relate only to the sample tested. Full results are contained in the relevant works order file.

RINGWOOD

Holly Grove Farm, Verwood Road, Ashley, Ringwood, Hants, BH24 2DB, UK.

T +44 (0)1425 479979 F +44 (0)1425 480637 E test@tracglobal.com

www.tracglobal.com

EMISSIONS

<i>Test Standard</i>	<i>Class</i>	<i>Pass</i>	<i>Pass Within Limits of Uncertainty</i>
EN 55022:1998 +A1:2000 +A2:2003 – Radiated Electromagnetic Emissions	A <input checked="" type="checkbox"/> B <input type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
EN 55022:1998 +A1:2000 +A2:2003 – Conducted Electromagnetic Emissions	A <input type="checkbox"/> B <input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>

<i>Test Standard</i>	<i>Class</i>	<i>Pass</i>	<i>N/A</i>
EN 61000-3-2:2006 – Harmonic Current Emissions	A <input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
EN 61000-3-3:1995 +A1:2001 +A2:2005 – Voltage Fluctuations and Flicker		<input checked="" type="checkbox"/>	

IMMUNITY

<i>Basic Test Standard</i>	<i>Applicable</i>	<i>Performance Criteria Required</i>	<i>Performance Criteria Achieved</i>
EN 61000-4-2:1995 +A1:1998 +A2:2001 – Electrostatic Discharge	<input checked="" type="checkbox"/>	B	A
EN 61000-4-3:2006 – Radiated Immunity	<input checked="" type="checkbox"/>	A	A
EN 61000-4-4:2004 – Electrical Fast Transients	<input checked="" type="checkbox"/>	B	A
EN 61000-4-5:2006 – Voltage Surge	<input checked="" type="checkbox"/>	B	A
EN 61000-4-6:2007 – Conducted Radio Frequency Immunity	<input checked="" type="checkbox"/>	A	A
EN 61000-4-11:2004 – Voltage Dips and Short Interruptions	<input checked="" type="checkbox"/>	B/C	B/B