

Certificate of Test

CERTIFICATE No: TES-001547-02
ISSUE No: 1
COPY No: 1
DATE: 10th August 2010

PURPOSE OF TEST: EMC EMISSIONS AND IMMUNITY TESTING
TRaC PROJECT ID: TES-001547-00
TEST SPECIFICATION(S): EN61326-1:2006 - Class A Emissions
Industrial Immunity Requirements
IEC61000-4-2 ESD
IEC61000-4-3 Radiated Immunity
IEC61000-4-4 Fast Transient Burst
IEC61000-4-5 Surges
TEST(S) CARRIED OUT IEC61000-4-6 Conducted Immunity
IEC61000-4-11 Voltage Dips and Interruptions
EN55011 Class A Conducted Emissions
EN55011 Class A Radiated Emissions
DEVIATIONS FROM SPECIFICATION(S): NIL
EQUIPMENT UNDER TEST (EUT): Bioquell R-30
DESCRIPTION OF EUT: Aeration device part of a bio decontamination system
MODEL / TYPE / VERSION No: TD066-0000-230V
CATALOGUE / SAMPLE No: None Supplied
EQUIPMENT SERIAL No: 201008R30001
TEST RESULT: MEASURED AS COMPLIANT :
(Measurement uncertainty as per RF109 current issue)

The results herein relate only to the sample tested. Full results are contained in the relevant works order file.

Certificate of Test

CERTIFICATE No: TES-001547-02
ISSUE No: 1
COPY No: 1
DATE: 10th August 2010

MANUFACTURER / AGENT: BIOQUELL UK LTD
52 ROYCE CLOSE
WEST PORTWAY INDUSTRIAL ESTATE
ANDOVER
HAMPSHIRE
SP10 3TS

ORDER No / CLIENT REFERENCE: STD235149
DATE OF TEST: 4TH TO 10TH AUGUST 2010

TESTED BY:



TRaC
regulatory and compliance



F. BARKAS

APPROVED BY:



TRaC
regulatory and compliance



S. YOUNGMAN

General Manager, TRaC EMC & Safety Ltd Ringwood

The results herein relate only to the sample tested. Full results are contained in the relevant works order file.

RINGWOOD

Holly Grove Farm, Verwood Road, Ashley, Ringwood, Hants, BH24 2DB, UK.
T +44 (0)1425 479979 F +44 (0)1425 480637 E test@tracglobal.com
www.tracglobal.com